


Certificate of Analysis

| Product Name | | |
|-----------------------------------|--------------|-----------------|
| Psychiatry Olanzapine Reagent Kit | | |
| Catalog Number | Lot Number | |
| C82915 | 5109-A | |
| Manufacture Date | Release Date | Expiration Date |
| 2025-07-08 | 2025-07-08 | 2026-12-01 |

| Document Reference | Version | Parameter Tested | Specification | Result |
|--------------------|---------|-------------------------|------------------|--------|
| WI070 | 17 | Low Control Recovery | $\leq \pm 15\%$ | Pass |
| | | Medium Control Recovery | $\leq \pm 15\%$ | Pass |
| | | High Control Recovery | $\leq \pm 15\%$ | Pass |
| | | Precision (%CV) | $\leq 10\%$ | Pass |
| | | Fill Volume | ≥ 100 Tests | Pass |

This certifies that the Saladax Biomedical, Inc. manufactured item listed above has been produced, controlled, and has met all applicable specifications for release at the Bethlehem, PA facility in accordance with the procedures and requirements of the Saladax Quality Management System. Products are manufactured/assembled and tested in compliance to ISO 13485:2016 requirements.

| | |
|--------------------------------|---|
| Manufacturer Name and Address: | Saladax Biomedical, Inc. 116 Research Dr. Bethlehem, PA 18015 USA |
|--------------------------------|---|

| | |
|---|-------------|
| Released By (Quality Management): | Date: |
|  | 10 Jul 2025 |

Distributed by: Beckman Coulter, Inc. 250 S. Kraemer Blvd. Brea, CA 92821